

SCHNEIDER ELECTRIC INDIA PRIVATE LIMITED

MPS, Gate No 1, Plot No. 324-330, KIADB Industrial Area, Hebbal-Hootagalli, Mysuru - 576 013.

TEST CERTIFICATE

ITEM : Single Phase Static Watthour Meter
METER SL NO. : A9206655
RATING : 5-30A, 240V, 50Hz
METER CLASS : 1.0
CONSTANT : 3200 Impulses per kWh
STANDARD : IS 13779

The Meter was tested & has passed following routine tests as per IS: 13779 at reference conditions:

	TEST	RESULT
Q1	No-Load Conditions	Passed
Q2	AC Voltage Test	Passed
Q3	Insulation Resistance Test	Passed
Q4	Starting Test	Passed
Q5	Meter Accuracy Test	Passed Meter Accuracy Observed within $\pm 1\%$

Initial Reading

00001

CERTIFIED

Signature

Date: 26/08/2024

Meter Tester Gr - I
MTI Laboratory
MePDCL, Lumjingshal
Shillong

Executive Engineer
MTI Division
MePDCL, Shillong